

Especificaciones Técnicas

NO.	GAS	ALCANCE	PRECISION DE DETECCIÓN	90% DE TIEMPO DE RESPUESTA	LECTURA MINIMA	PUNTO DE AJUSTE DE ALARMA ALTA / BAJO
1	PH3	0-2000ppm	± 5% (FS)	≤25s	0.01ppm	0.3 / 5ppm
2	O2	0-30% vol	± 5%	≤15s	0,10%	19,5% / 23,0%
3	EX	0-100% LEL	± 5% (FS)	≤5s	1% LEL	25.0% / 50.0% LEL
4 4	CO	0-1000ppm	± 5% (FS)	≤25s	1ppm	50ppm / 100ppm
5 5	CO2	0-5000ppm	± 5% (FS)	≤25s	1ppm	1000ppm / 2000ppm
6 6	H2S	0-100ppm	± 5% (FS)	≤30s	1ppm	10ppm / 20ppm
7 7	NO2	0-20ppm	± 5% (FS)	≤25s	0.1ppm	3ppm / 10ppm
8	NO	0-250ppm	± 5% (FS)	≤60s	1ppm	25ppm / 50ppm
9 9	SO2	0-100ppm	± 5% (FS)	≤30s	0.1ppm	2ppm / 10ppm
10	CL2	0-20ppm	± 5% (FS)	≤30s	0.1ppm	1ppm / 10ppm
11	NH3	0-100ppm	± 5% (FS)	≤50s	1ppm	25ppm / 50ppm
12	H2	0-1000ppm	± 5% (FS)	≤60s	1ppm	50ppm / 10ppm
13	HCN	0-50ppm	± 5% (FS)	≤200s	0.1ppm	10ppm / 20ppm
14	HCL	0-20ppm	± 5% (FS)	≤60s	0.1ppm	5ppm / 10ppm
15	CH4	0-100% LEL	± 5% (FS)	≤5s	1% LEL	25.0% / 50.0% LEL